



Quality Certificate

VWR SYRINGE FILTER 25MM

This document certifies that the specified product meets all manufacturing specifications and criteria for release. Materials of construction have been evaluated in accordance with United States Pharmacopeia (USP) Biological Reactivity test In Vivo <88> (USP Class VI-121° C Plastic Tests) to ensure biological safety of the material. Pall Corporation also certifies that this product is non-pyrogenic as determined by Limulus Amebocyte Lysate method and is sterilized by Gamma Irradiation at a validated dosage range of 25-40 KiloGrays (2.5-4.0 mRads).

Product Number:	2 8 1 4 5 - 4 7 7
VWR North American Catalog No.:	28145-477
VWR European Article No.:	514-0061
Description:	25 mm Filter w/0.2 µm Cellulose Acetate
Lot Number:	1 1 3 4 8 7 0
Shipment Information:	
Manufacture Date:	November 2010
Sterilization Date:	December 2010
Expiration Date:	November 2013
Materials of Construction:	
Filter Media:	Cellulose Acetate Hydrophilic
Housing:	Modified Acrylic
Specifications:	
Effective Filtration Area:	2.8 cm ²
Pyrogenicity:	< 0.5 EU/mL
Dosage Range:	25 - 40 KiloGrays

Signed:

Ken Crossley
Manager
Quality Assurance

February , 2011