



Quality Certificate

VWR SYRINGE FILTERS 25 MM

This document certifies that the specified product meets or exceeds our specifications as outlined in our currently published literature. Materials of construction have been evaluated in accordance with United States Pharmacopeia (USP) Biological Reactivity Test, In Vivo <88> (USP Class VI-121 °C Plastic Tests) to ensure biological safety of the material. Pall Corporation also certifies that this product is non-pyrogenic as determined by Limulus Amebocyte Lysate method and is sterilized by Gamma Irradiation.

Product Number:	
VWR European Article No.:	514-0073
Description:	PES Sterile Syringe Filter, 25 mm
Pore Size:	0.2 µm
Lot Number:	189034
Shipment Information:	
Manufactured Date:	August 2008
Sterilization Date:	September 2008
Materials of Construction:	
Filter Media:	Polyethersulfone Membrane
Housing:	Modified Acrylic
Specifications:	
Effective Filtration Area:	2.8 cm ²
Dosage Range :	25 – 40 KiloGrays
Maximum Operating Temperature:	55 °C (131 °F)

Signed:

Ken Crossley
Manager
Quality Assurance

December, 2010